



Certificate of Analysis

Description: GMO Chek Soy Elisa Kit
Part Number: 7099999
Lot Number: 09E1059
Expiration Date: September 2010
Storage Condition: Refrigerated

Quality Control Testing

Test	Spec	Results	Pass/Fail
0 ng/mL Blank	< 0.12	0.06	Pass
0% Protein Isolate	< 0.12	0.08	Pass
2.5% Protein Isolate	1.0 – 2.0	1.1	Pass
Correlation of Protein Isolate Standard (Quadratic Fit)	> 0.96	1.00	Pass
Quality Control Sample 1 (0.1%)	≤ 0.20%	0.09%	Pass
Quality Control Sample 2 (0.7%)	0.56% - 0.84%	0.69%	Pass

The values above are the mean of all Quality Control results obtained for this lot. Results are affected by temperature, time, and pipette accuracy. Test performed in a controlled temperature setting with calibrated instrumentation. Individual results may vary from results listed above.

Constance Viering
Director of Kit Operations

Insert
Approval
Stamp Here

7/28/2009

Alan McQuillin
Director of Quality