



## Certificate of Analysis

---

**Description:** Trait Chek RUR Lateral Flow Test Strip  
**Part Number:** 3000082  
**Lot Number:** 09H1040x  
**Expiration Date:** February 2011

**Storage Condition:** Room Temperature

### Quality Control Testing

Test	Results	Specification	Pass/Fail
Control line visible on all strips	Visible on all strips	Visible on all strips	Pass
Negative Control @ 5 Minutes (Non-GMO Soybean)	No Test Line present	No Test Line present	Pass
Positive Control @ 5 Minutes (RUR Soybean)	Test Line visible	Test Line visible	Pass
Negative Control @ 60 Minutes (Non-GMO Soybean)	No Test Line present	No Test Line present	Pass

The values above are the mean of all Quality Control results obtained for this lot.

Constance Viering  
Director of Kit Operations

Insert  
Approval  
Stamp Here

9/14/2009

Alan McQuillin  
Director of Quality