



Certificate of Analysis

Description: RapidChek ® Base Enrichment Media for Listeria

Part Number: 7000172, 7000177

Lot Number: K08-25-2

Expiration Date: December 2012

Storage Condition: Room Temperature

Physical and Chemical Characteristics

Appearance (Dry): Homogeneous, free flowing, beige

Color/Clarity (Reconstituted): Yellow/clear

pH (5.3% solution at 25°C: 7.2 +/- 0.2): 7.22

Quality Control Testing

Test	Results	Specification	Pass/Fail
L.monocytogenes ATCC 7644 Control Organism	Good Growth	Good Growth	Pass
L.monocytogenes ATCC 19115 Control Organism	Good Growth	Good Growth	Pass
L.monocytogenes ATCC 19112 Control Organism	Good Growth	Good Growth	Pass
S.aureus ATCC 25923	No Growth	No Growth	Pass
P.aeruginosa ATCC 27853	No Growth	No Growth	Pass
E. coli ATCC 25922	No Growth	No Growth	Pass

Constance Viering
Director of Kit Operations

Insert
Approval
Stamp Here

4/23/2009

Alan McQuillin
Director of Quality