



## Certificate of Analysis

---

**Product Description: RapidChek® Listeria Test Strip**

**Part Numbers: 3000019**

**Lot Number: 09A1055**

**Expiration Date: July 2010**

**Storage Condition: Room Temperature**

**Quality Control Testing**

Test	Results	Specification	Pass/Fail
Control line visible on all strips	Visible on all strips	Visible on all strips	Pass
Negative Control @ 10 Minutes	No Test Line present	No Test Line present	Pass
Negative Control @ 60 Minutes	No Test Line present	No Test Line present	Pass
Positive Control @ 10 Minutes	Test Line visible	Test Line visible	Pass

The values above are the mean of all Quality Control results obtained for this lot

Constance Viering  
Director of Kit Operations

Insert  
Approval  
Stamp Here

2/17/2009

Alan McQuillin  
Director of Quality