



Certificate of Analysis

Description: Rapid Chek Salmonella Primary Enrichment Medium

Part Number: 7000184, 7000192

Lot Number: B09-45

Expiration Date: March 2013

Storage Condition: Room Temperature

Physical and Chemical Tests

Appearance (Dry): Homogeneous, free flowing w/ soft lumps, light beige

Color/Clarity (Reconstituted): Light yellow/ clear

pH (2.0% solution at 25°C): 7.05

Quality Control Testing

Test	Results	Specification	Pass/Fail
Salmonella enteritidis ATCC 13076	Good Growth	High recovery @ 4-6hrs, heavy growth in test tubes @ 8-24hrs	Pass
Salmonella Montevideo SDI ARS32	Good Growth	High recovery @ 4-6hrs, heavy growth in test tubes @ 8-24hrs	Pass
Salmonella typhimurium ATCC 14028	Good Growth	High recovery @ 4-6hrs, heavy growth in test tubes @ 8-24hrs	Pass

** The following results were obtained from type cultures after incubation at a temperature of 42° C and observed after 4-24 hours.

Constance Viering
Director of Kit Operations

Insert
Approval
Stamp Here

4/22/2009

Alan McQuillin
Director of Quality