

Description: Envirogard Microcystins Plate**Part Number:** 7540000**Lot Number:** 11E1218**Expiration Date:** October 2012**Storage Condition:** Refrigerated, 2 to 8 °C**Quality Control Testing**

Test	Results	Specification	Pass/Fail
Correlation Coefficient	0.98	≥ 0.98	Pass
1.0 ppb Microcystins LR Standard Recovery Value*	1.0	0.9 – 1.3 ppb	Pass

The values above are the mean of all Quality Control results obtained for this lot.

* prepared from Microcystins LR standard, Sigma part #: 33983; Oekanal®, 10 µg/mL in Methanol (Reidel-de-Haen)



Constance Viering
Director of Kit Operations



Alan McQuillin
Director of Quality