



## Certificate of Analysis

---

**Product Description:** Envirogard Microcystins Quantitube Kit  
**Part Number:** 7700000  
**Lot Number:** 09D1037  
**Expiration Date:** October, 2010  
**Storage Condition:** Refrigerated

### Quality Control Testing

Test	Results	Specification	Pass/Fail
Mean Bo (Negative Control OD)	2.0	1.0 – 2.0	Pass
Correlation Coefficient (R <sup>2</sup> )	1.00	≥ 0.98	Pass

\* The values above are the mean of all Quality Control results obtained for this lot. \*

Constance Viering  
Director of Kit Operations

Insert  
Approval  
Stamp Here

4/22/2009

Alan McQuillin  
Director of Quality