



**Description: *Envirogard Microcystins Quantitube Kit***

**Part Number:** 7700000

**Lot Number:** 10K1008R

**Expiration Date:** February 2012

**Storage Condition:** Refrigerated

**Quality Control Testing**

Test	Results	Specification	Pass/Fail
Mean Bo (Negative Control OD)	1.5	1.0 - 2.0	Pass
Correlation Coefficient (R)	1.00	≥ 0.98	Pass

\* The values above are the mean of all Quality Control results obtained for this lot. \*

Constance Viering  
Director of Kit Operations



Alan McQuillin  
Director of Quality