

Description: *Envirogard Microcystins Quantitube Kit*

Part Number: 7700000

Lot Number: 11C1086

Expiration Date: July 2012

Storage Condition: Refrigerated

Quality Control Testing

Test	Results	Specification	Pass/Fail
Mean Bo	1.7	1.0 – 2.0	Pass
Correlation Coefficient (R)	0.99	≥ 0.98	Pass

The values above are the mean of all Quality Control results obtained for this lot.



Constance Viering
Director of Kit Operations



Alan McQuillin
Director of Quality