

**Description:** *Envirogard Microcystins Quantitube Kit*

**Part Number:** 7700000

**Lot Number:** 11C1086A

**Expiration Date:** July 2012

**Storage Condition:** Refrigerated

**Quality Control Testing**

Test	Results	Specification	Pass/Fail
Mean Bo	1.7	1.0 – 2.0	Pass
Correlation Coefficient (R)	0.99	≥ 0.98	Pass

The values above are the mean of all Quality Control results obtained for this lot.



Constance Viering  
Director of Kit Operations



Alan McQuillin  
Director of Quality