



Description: *Microtox*[®] 2% Diluent-1Liter

Part Number: AZF686011

Lot Number: 11L1502

Expiration Date: September 2014

Storage Condition: Room Temperature

Quality Control Testing

Test	Results	Specification	Pass/Fail
Comparison % - 5 Min	3.4%	$\leq \pm 5\%$	Pass
Comparison % - 15 Min	2.4%	$\leq \pm 5\%$	Pass
Conductivity result	2.23%	1.68% - 2.52%	Pass

The values above are the mean of all Quality Control results obtained for this lot. All comparison results are obtained using the Quality Control Procedure with an approved lot of Microtox 2% Diluent on the Model 500.

Reviewed by:

11/28/11