



Certificate of Analysis

Description: Microtox Acute Reagent
Part Number: AZF686018A
Lot Number: 09A1032(A)
Expiration Date: January 2011
Storage Condition: Frozen, -20 °C

Quality Control Testing

Test	Results	Specification	Pass/Fail
Phenol	EC50 (5 min) = 18.3 mg/L	13.0-26.0 mg/L	Pass
Zinc Sulfate	EC50 (15 min) = 8.5 mg/L	3.0-10.0 mg/L	Pass
Zinc ++ (ion) *Calculated	EC50 (15 min) = 1.9 mg/L	0.6-2.2 mg/L	Pass
DeltaTox Adjusted Light Reading	89	≥ 30	Pass

The values above are the mean of all Quality Control results obtained for this lot. All results are obtained using the Quality Control Procedure with an approved lot of Microtox 2% Diluent and Reconstitution Solution on the Model 500 and DeltaTox Analyzers.

Constance Viering
Director of Kit Operations

Insert
Approval
Stamp Here

3/6/2009

Alan McQuillin
Director of Quality